

## Certificate of Traceability

## **Manufacturer Certification**

This certifies that <u>TaiDoc Technology Corporation</u>, using <u>Immunoassay</u>

on TD-4611 has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **European Reference Laboratory ESRL#13** 

The system evaluated was:

| Instrument:  | Calibrator Lot:                                |
|--------------|--|
| TD-4611      | 704101-122, 704101-122, 704101-122, 704101-122 |
| Reagent Lot: | Calibrator Assigned Values:                    |
| 191223       | 5.5%, 7.9%, 10.7%, 15.5%                       |

Randie R Little PhD.

Date of Certification: January 1, 2020

Certification Expires: January 1, 2021

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor