

TD-4280

BLOOD GLUCOSE MONITORING SYSTEM

5 Sec, 0.5 μ L

Strip-Ejection

Voice Guidance

FEATURES

- Before / After meal (AC/PC) presetable
- Voice guidance for measuring result
- Ketone warning
- Display calendar and clock when meter standby



SPECIFICATIONS

Enzyme Type	GDH-FAD
Sample Size	0.5 μ L
Reaction Time	5 seconds
Measurement Range	20 ~ 600 mg/dL (1.1 ~ 33.3 mmol/L)
Hematocrit Range	35% - 60%
Precision	SD < 5 mg/dL (0.278 mmol/L) at < 100 mg/dL (5.55 mmol/L); CV < 5% at 100 mg/dL (5.55 mmol/L)
Accuracy	\pm 15 mg/dL if \leq 100 mg/dL; \pm 15% if >100 mg/dL
Ketone Warning	Yes
Communication	RS232
Power Source	AAA x 2
Memory Capacity	450 sets
Day Average	7, 14, 21, 28, 60, 90 days
Daily Alarm	4 daily alarms
Dimension	95 (L) x 52 (W) x 20 (H) mm
Weight	48.8 g (without battery)
Operating Condition	10°C (50°F) ~ 40°C (104°F)
Storage Condition	-20°C (-4°F) ~ +60°C (140°F) (Meter) 2°C (35.6°F) ~ 32°C (89.6°F) (Strip)

SHIPPING INFO

Units per Carton	40
Gross Weight per Carton	12.8 kg
Carton Size	49.8 (L) x 35.8 (W) x 36 (H) cm



www.taidoc.com

TaiDoc Technology Corp.

B1-7F., No.127, Wugong 2nd Rd., Wugu Dist.,
New Taipei City 24888, Taiwan
Tel : +886-2-6625-8188 Fax : +886-2-6625-0288